



About Touch Bionics

Touch Bionics is a provider of world-leading prosthetic technologies and supporting services designed to ensure the best possible outcomes for people with upper limb deficiencies. Our advanced products include myoelectric prosthetic hand and prosthetic finger solutions, as well as highly realistic passive silicone prostheses that match the natural appearance of the wearer. We are committed to helping ensure that our patients and customers have the best possible experience with our products by supporting them with training and a full range of clinical services.

North America: +1 855 MY iLIMB (694 5462)
customersupportus@touchbionics.com

Germany/Europe: +49 6221 357 9060
customersupport@touchbionics.de

France: 0805 110478
customersupport@touchbionics.fr

United Kingdom/International: +44 1506 438 556
customersupport@touchbionics.com

www.touchbionics.com



i-limb™

Quick Reference Guide



i-limb™ Quick Reference Guide

Downloading biosim & system requirements

To register for biosim™, please go to www.touchbionics.com/biosimdownload

- Windows XP, Windows Vista, Windows 7, Windows 8, or Mac OS X Lion (10.7 and higher)
- Microsoft .NET Framework v3.5
- USB port for Bluetooth® connector (available for order)
- Administrative rights to install the software and connector

Downloading the my i-limb™ mobile app

my i-limb™ mobile control app is available on the Apple® App store for download and is compatible with iPhone® (4s or newer), iPad® (4th generation or newer), iPad® mini, iPod (5 or newer).

Product Manuals

Please refer to www.touchbionics.com/downloads/document-library to ensure you have the latest copy of the product manual. To request a printed manual, please contact Customer Support.

Battery Charging

To ensure compatibility and optimum performance use ONLY the Touch Bionics charger as supplied with your device.



When charging the battery, remove the prosthesis from the residual limb and switch the hand to the OFF position.



Battery should be replaced annually for best performance and optimum battery life.



i-limb™ Power Pack users should refer to the specific device manual for charging instructions.

Safety Guidelines

Please read these guidelines and information before using the i-limb™ product. Failure to do so may lead to risk of product malfunction or even a potential hazard to your health, and may lead to an invalidation of the warranty.

- | | |
|-------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|-----------------------------------------------------------------------------------------------------------------------|
| Do not use without an approved cover | Do not expose to high temperatures |
| Do not use under water | Do not expose to flames |
| Do not use to operate heavy/industrial machinery | Do not use or expose to explosive atmospheres |
| Do not use with machinery with moving parts that may cause personal injury or damage | Do not disassemble componentry or modify in any way |
| Users must comply with local regulations on the operation of automobiles, aircraft, sailing vessels of any kind and any other motorized vehicle or device (Please refer to the user manual for Touch Bionics' Driving of Motor Vehicles Position Statement) | Maintenance, repairs and upgrades may only be performed by qualified Touch Bionics technicians and technical partners |
| Do not use for extreme activities that may cause injury to a natural hand e.g. rock climbing | Do not use with a damaged cover |
| Do not expose to excessive moisture, liquid, dust, vibration or shock | Damaged covers must be replaced or repaired by a qualified Touch Bionics technician or technical partner |
| | Only approved Touch Bionics accessories and tooling may be used with i-limb™ devices |

See www.touchbionics.com/downloads/document-library for further General Safety, EMC and Electrical information for i-limb™ products.